


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL080320\
 Data File : VL035314.D
 Acq On : 3 Aug 2020 10:31
 Operator : SY/AP
 Sample : BFB
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL080320AIR.M
 Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Wed Jul 29 07:14:29 2020
 Last Update : Wed Jul 29 07:14:29 2020



AutoFind: Scans 3625, 3626, 3627; Background Corrected with Scan 3603

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	24.6	197492	PASS
75	95	30	66	60.1	481642	PASS
95	95	100	100	100.0	801856	PASS
96	95	5	9	6.1	48741	PASS
173	174	0.00	2	1.0	5199	PASS
174	95	50	120	66.3	531648	PASS
175	174	4	9	7.0	37125	PASS
176	174	93	101	96.2	511296	PASS
177	176	5	9	6.9	35421	PASS