

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL080524\
 Data File : VL041493.D
 Acq On : 5 Aug 2024 9:07
 Operator : SY/MD
 Sample : VL0805ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0805ABL01

Quant Time: Aug 06 01:35:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Aug 06 01:32:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.189	49	1177454	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.645	114	3167373	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.459	117	2871846	10.000	ppbv	# 0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.770	95	2211478	10.006	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.100%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL080524\
Data File : VL041493.D
Acq On : 5 Aug 2024 9:07
Operator : SY/MD
Sample : VL0805ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0805ABL01

Quant Time: Aug 06 01:35:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Tue Aug 06 01:32:58 2024
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 5.189 min Scan# 804
Delta R.T. -0.000 min
Lab File: VL041493.D
Acq: 5 Aug 2024 9:07




Instrument :
MSVOA_L
ClientSampleId :
VL0805ABL01


Tgt Ion: 49 Resp: 1177454
Ion Ratio Lower Upper
49 100
128 47.6 23.8 71.4
130 60.5 30.8 92.4



#33
1,4-Difluorobenzene
Concen: 10.000 ppbv
RT: 6.645 min Scan# 1257
Delta R.T. -0.003 min
Lab File: VL041493.D
Acq: 5 Aug 2024 9:07

Tgt Ion: 114 Resp: 3167373
Ion Ratio Lower Upper
114 100
63 24.8 20.3 30.5
88 17.9 14.3 21.5









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 11.459 min Scan# 21
Delta R.T. -0.003 min
Lab File: VL041493.D
Acq: 5 Aug 2024 9:07

Instrument :
MSVOA_L
ClientSampleId :
VL0805ABL01

Tgt Ion:117 Resp: 2871846
Ion Ratio Lower Upper
117 100
82 64.4 51.0 76.6
119 33.7 45.2 67.8#



#68
1-Bromo-4-Fluorobenzene
Concen: 10.006 ppbv
RT: 13.770 min Scan# 3473
Delta R.T. 0.003 min
Lab File: VL041493.D
Acq: 5 Aug 2024 9:07

Tgt Ion: 95 Resp: 2211478
Ion Ratio Lower Upper
95 100
174 71.3 56.3 84.5
175 5.0 4.0 6.0

