

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL080824\
 Data File : VL041523.D
 Acq On : 8 Aug 2024 22:01
 Operator : SY/MD
 Sample : VIBLK
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 08/09/2024
 Supervised By :Mahesh Dadoda 08/09/2024

Quant Time: Aug 09 06:00:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 9 18:33:45 2024
 QLast Update : Tue Aug 06 01:32:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.202	49	995236	10.000	ppbv	0.01
33) 1,4-Difluorobenzene	6.658	114	2770306	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.465	117	2260332	10.000	ppbv	# 0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	13.793	95	1743030m	10.020	ppbv	0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL080824\
Data File : VL041523.D
Acq On : 8 Aug 2024 22:01
Operator : SY/MD
Sample : VIBLK
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 08/09/2024
Supervised By :Mahesh Dadoda 08/09/2024

Quant Time: Aug 09 06:00:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Tue Aug 06 01:32:58 2024
Response via : Initial Calibration

