

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL081020\
 Data File : VL035375.D
 Acq On : 10 Aug 2020 15:38
 Operator : SY/AP
 Sample : QC080320 CAN 10439
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC080320 CAN 10439

Quant Time: Aug 11 05:00:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL080520AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Aug 05 13:04:46 2020
 QLast Update : Wed Aug 05 13:04:46 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1187816	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	2934160	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.11	117	2831240	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.45	95	2387612	10.11	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	3.83	43	65209	0.387	ppbv #	68
20) Methylene Chloride	4.32	84	163349	2.375	ppbv	96
26) Hexane	5.67	57	86641	0.779	ppbv #	39

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL081020\
Data File : VL035375.D
Acq On : 10 Aug 2020 15:38
Operator : SY/AP
Sample : OC080320 CAN 10439
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC080320 CAN 10439

Quant Time: Aug 11 05:00:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL080520AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Aug 05 13:04:46 2020
QLast Update : Wed Aug 05 13:04:46 2020
Response via : Initial Calibration

