

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL081020\
 Data File : VL035377.D
 Acq On : 11 Aug 2020 7:20
 Operator : SY/AP
 Sample : L3554-04DL 40X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 4-UNDER-GROUND DL

Manual Integrations
 APPROVED

MMDadoda
 8/12/2020 3:58:25 PM

Quant Time: Aug 11 20:41:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL080520AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Aug 0
 QLast Update : Wed Aug 05 13:04:46 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1243149	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	3110988	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.11	117	2849785	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.45	95	2401413	10.10	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.00%

Target Compounds					Ovalue
9) Propene	2.97	41	41242	0.816 ppbv	90
11) Trichlorofluoromethane	3.91	101	160278	0.748 ppbv	93
15) Acetone	3.81	43	1316156	7.455 ppbv	96
34) 2-Butanone	5.23	43	192648m	1.258 ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL081020\
Data File : VL035377.D
Acq On : 11 Aug 2020 7:20
Operator : SY/AP
Sample : L3554-04DL 40X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
4-UNDER-GROUNDL

Manual Integrations
APPROVED

MMDadoda
8/12/2020 3:58:25 PM

Quant Time: Aug 11 20:41:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL080520AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Aug
QLast Update : Wed Aug 05 13:04:46 2020
Response via : Initial Calibration

