

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL081220\
 Data File : VL035388.D
 Acq On : 12 Aug 2020 18:21
 Operator : SY/AP
 Sample : VIBLK
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Manual Integrations
 APPROVED

MMDadoda
 8/17/2020 5:43:59 PM

Quant Time: Aug 13 04:02:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL081220AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Aug 13 04:02:52 2020
 QLast Update : Wed Aug 12 16:45:58 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1466611	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	3369101	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.10	117	3185805	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.44	95	2664677	10.42	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	104.20%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.26	62	4066m	0.041	ppbv	
20) Methylene Chloride	4.34	84	11745	0.168	ppbv	93
32) 1,1,1-Trichloroethane	6.51	97	9311m	0.035	ppbv	
35) Carbon Tetrachloride	7.01	117	9080	0.038	ppbv #	59
38) Trichloroethene	7.83	130	6354m	0.048	ppbv	
52) Tetrachloroethene	11.24	164	5696m	0.048	ppbv	
80) Naphthalene,2-methyl-	24.99	142	36624	0.130	ppbv	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL081220\
Data File : VL035388.D
Acq On : 12 Aug 2020 18:21
Operator : SY/AP
Sample : VIBLK
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Manual Integrations
APPROVED

MMDadoda
8/17/2020 5:43:59 PM

Quant Time: Aug 13 04:02:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL081220AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Aug
QLast Update : Wed Aug 12 16:45:58 2020
Response via : Initial Calibration

