

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL081320\
 Data File : VL035402.D
 Acq On : 13 Aug 2020 15:33
 Operator : SY/AP
 Sample : L3510-01
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 E053-1100

Quant Time: Aug 15 15:01:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL081220AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Wed Aug 12 20:32:44 2020
 QLast Update : Wed Aug 12 20:32:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1346198	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	3006329	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.10	117	2976005	10.00	ppbv	-0.01

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.43	95	2854488	11.88	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	118.80%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
6) Bromomethane	3.47	94	1866599	38.938	ppbv	99
18) 1,1-Dichloroethene	4.28	96	1799562	29.358	ppbv	94
20) Methylene Chloride	4.34	84	854961	13.353	ppbv	98
29) Chloroform	5.75	83	4628347	21.178	ppbv	95
38) Trichloroethene	7.84	130	3298472	29.196	ppbv	97
42) Bromodichloromethane	7.80	83	7510348	33.318	ppbv #	91
47) 1,1,2-Trichloroethane	9.48	97	3063609	26.460	ppbv	98
48) Dibromochloromethane	10.32	129	4790924	25.889	ppbv	99
58) Ethyl Benzene	12.73	91	8404822	17.869	ppbv	89
76) 1,4-Dichlorobenzene	17.16	146	7191672	28.063	ppbv	95
77) 1,2-Dichlorobenzene	17.84	146	6988314	27.782	ppbv	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL081320\
Data File : VL035402.D
Acq On : 13 Aug 2020 15:33
Operator : SY/AP
Sample : L3510-01
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
E053-1100

Quant Time: Aug 15 15:01:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL081220AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Aug 12 20:32:44 2020
QLast Update : Wed Aug 12 20:32:44 2020
Response via : Initial Calibration

