

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL081320\
 Data File : VL035406.D
 Acq On : 13 Aug 2020 18:24
 Operator : SY/AP
 Sample : L3510-01DL 2X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 E053-1100DL

Quant Time: Aug 15 15:24:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL081220AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Aug 12 20:32:44 2020
 QLast Update : Wed Aug 12 20:32:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1308538	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	2946924	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.10	117	3106341	10.00	ppbv	-0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.43	95	3200174	12.76	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	127.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
6) Bromomethane	3.47	94	965304	20.716	ppbv	96
18) 1,1-Dichloroethene	4.29	96	923431	15.498	ppbv	98
20) Methylene Chloride	4.34	84	327956	5.270	ppbv	90
29) Chloroform	5.75	83	2558205	12.042	ppbv	97
38) Trichloroethene	7.84	130	1809850	16.343	ppbv	98
42) Bromodichloromethane	7.80	83	4532326	20.512	ppbv	97
47) 1,1,2-Trichloroethane	9.48	97	1636799	14.422	ppbv	97
48) Dibromochloromethane	10.31	129	2584144	14.245	ppbv	100
58) Ethyl Benzene	12.73	91	4824472	9.826	ppbv	99
76) 1,4-Dichlorobenzene	17.16	146	3983624	14.892	ppbv	97
77) 1,2-Dichlorobenzene	17.83	146	3845622	14.647	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL081320\
Data File : VL035406.D
Acq On : 13 Aug 2020 18:24
Operator : SY/AP
Sample : L3510-01DL 2X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
E053-1100DL

Quant Time: Aug 15 15:24:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL081220AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Aug 12 20:32:44 2020
QLast Update : Wed Aug 12 20:32:44 2020
Response via : Initial Calibration

