

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL081321\
 Data File : VL037447.D
 Acq On : 13 Aug 2021 14:00
 Operator : SY/AP
 Sample : M1458-14 0.03 LOQ
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 LOQ-AIR-01-QT1-2021

Manual Integrations
 APPROVED

MMDadoda
 8/16/2021 5:04:02 PM

Quant Time: Aug 16 02:59:43 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL081221AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 13 2021
 QLast Update : Fri Aug 13 01:37:35 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1233959	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.17	114	3755620	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.08	117	3263327	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.41	95	2301753	9.83	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.30%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.25	62	4045m	0.042	ppbv	
32) 1,1,1-Trichloroethane	6.50	97	7882m	0.034	ppbv	
35) Carbon Tetrachloride	7.01	117	8550m	0.037	ppbv	
38) Trichloroethene	7.82	130	5167m	0.037	ppbv	
52) Tetrachloroethene	11.21	164	4755m	0.036	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.67	83	11955m	0.044	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL081321\
Data File : VL037447.D
Acq On : 13 Aug 2021 14:00
Operator : SY/AP
Sample : M1458-14 0.03 LOQ
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
LOQ-AIR-01-QT1-2021

Manual Integrations
APPROVED

MMDadoda
8/16/2021 5:04:02 PM

Quant Time: Aug 16 02:59:43 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL081221AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Aug 13 01:37:35 2021
Response via : Initial Calibration

