

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL081321\
 Data File : VL037449.D
 Acq On : 13 Aug 2021 15:13
 Operator : SY/AP
 Sample : M2262-14 0.03 LOQ
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 LOQ-AIR-02-QT2-2021

Manual Integrations
 APPROVED

MMDadoda
 8/16/2021 5:04:05 PM

Quant Time: Aug 16 02:50:24 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL081221AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 13 2021
 QLast Update : Fri Aug 13 01:37:35 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1179045	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.17	114	3693698	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.09	117	3178380	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.41	95	2326806	10.21	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.26	62	3299m	0.036	ppbv	
32) 1,1,1-Trichloroethane	6.52	97	8322m	0.038	ppbv	
35) Carbon Tetrachloride	7.02	117	8107m	0.036	ppbv	
38) Trichloroethene	7.83	130	5030m	0.037	ppbv	
52) Tetrachloroethene	11.23	164	4601m	0.035	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.66	83	11518m	0.043	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL081321\
Data File : VL037449.D
Acq On : 13 Aug 2021 15:13
Operator : SY/AP
Sample : M2262-14 0.03 LOQ
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
LOQ-AIR-02-QT2-2021

Manual Integrations
APPROVED

MMDadoda
8/16/2021 5:04:05 PM

Quant Time: Aug 16 02:50:24 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL081221AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Aug 13 01:37:35 2021
Response via : Initial Calibration

