

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL081618\
 Data File : VL032386.D
 Acq On : 16 Aug 2018 21:03
 Operator : SY/AP
 Sample : J4457-04DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SS02-MIVIDL

Manual Integrations
 APPROVED

apatel
 8/20/2018 3:18:30 PM

Quant Time: Aug 20 09:40:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL081518AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Aug 15 14:58:02 2018
 QLast Update : Wed Aug 15 14:58:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.82	49	1767422	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.35	114	3435267	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.31	117	3340801	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.64	95	2606164	10.23	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.30%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.09	41	54026m	0.77	ppbv	
13) Ethanol	3.70	45	63138	3.51	ppbv	97
15) Acetone	3.94	43	1159490	6.90	ppbv	95
26) Hexane	5.83	57	63842	0.40	ppbv #	46
29) Chloroform	5.90	83	182021	0.77	ppbv	99
34) 2-Butanone	5.40	43	71939	0.31	ppbv	100
35) Carbon Tetrachloride	7.19	117	27972m	0.14	ppbv	
52) Tetrachloroethene	11.44	164	573418	6.68	ppbv	95
53) Toluene	10.01	91	200221	0.63	ppbv	100
74) 1,2,4-Trimethylbenzene	16.99	105	44068m	0.12	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL081618\
Data File : VL032386.D
Acq On : 16 Aug 2018 21:03
Operator : SY/AP
Sample : J4457-04DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SS02-MIVIDL

Manual Integrations
APPROVED

apatel
8/20/2018 3:18:30 PM

Quant Time: Aug 20 09:40:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL081518AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Aug
QLast Update : Wed Aug 15 14:58:02 2018
Response via : Initial Calibration

