

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL081618\
 Data File : VL032388.D
 Acq On : 16 Aug 2018 22:23
 Operator : SY/AP
 Sample : J4457-06DL 40X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SG05-MIVIDL

Manual Integrations
 APPROVED

apatel
 8/22/2018 4:43:20 PM

Quant Time: Aug 22 16:37:24 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL081518AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Wed Aug 15 14:58:02 2018
 QLast Update : Wed Aug 15 14:58:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.82	49	1831944	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.35	114	3491372	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.31	117	3356281	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.64	95	2598739	10.15	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.50%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.07	41	141606	1.95	ppbv	94
13) Ethanol	3.70	45	17496m	0.94	ppbv	
15) Acetone	3.94	43	5913071	33.97	ppbv	100
19) Isopropyl Alcohol	4.07	45	5267648	44.94	ppbv #	99
26) Hexane	5.83	57	21861	0.13	ppbv #	39
34) 2-Butanone	5.39	43	55198	0.23	ppbv	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL081618\
Data File : VL032388.D
Acq On : 16 Aug 2018 22:23
Operator : SY/AP
Sample : J4457-06DL 40X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SG05-MIVIDL

Manual Integrations
APPROVED

apatel
8/22/2018 4:43:20 PM

Quant Time: Aug 22 16:37:24 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL081518AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Aug
QLast Update : Wed Aug 15 14:58:02 2018
Response via : Initial Calibration

