

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL081618\
 Data File : VL032389.D
 Acq On : 16 Aug 2018 23:04
 Operator : SY/AP
 Sample : J4457-07DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SG07-MIVIDL

Manual Integrations
 APPROVED

apatel
 8/22/2018 4:43:24 PM

Quant Time: Aug 22 16:37:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL081518AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Aug 1
 QLast Update : Wed Aug 15 14:58:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.81	49	1830245	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.35	114	3454840	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.30	117	3527531	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.64	95	2713636	10.08	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.80%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.07	41	298872	4.12	ppbv	91
10) Heptane	8.31	43	23250	0.11	ppbv	88
13) Ethanol	3.70	45	112422m	6.03	ppbv	
15) Acetone	3.94	43	1168971	6.72	ppbv	99
19) Isopropyl Alcohol	4.07	45	13357367	114.07	ppbv #	98
26) Hexane	5.83	57	69243	0.42	ppbv #	54
34) 2-Butanone	5.38	43	164674	0.70	ppbv	98
53) Toluene	10.01	91	161491	0.50	ppbv	93
59) m/p-Xylene	13.19	91	79260m	0.22	ppbv	
75) 1,3-Dichlorobenzene	17.23	146	68805	0.28	ppbv	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL081618\
Data File : VL032389.D
Acq On : 16 Aug 2018 23:04
Operator : SY/AP
Sample : J4457-07DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SG07-MIVIDL

Manual Integrations
APPROVED

apatel
8/22/2018 4:43:24 PM

Quant Time: Aug 22 16:37:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL081518AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Aug
QLast Update : Wed Aug 15 14:58:02 2018
Response via : Initial Calibration

