

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL081618\
 Data File : VL032390.D
 Acq On : 16 Aug 2018 23:44
 Operator : SY/AP
 Sample : J4457-08DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 DUP1-MIVIDL

Manual Integrations
 APPROVED

MMDadoda
 8/20/2018 3:11:53 PM

Quant Time: Aug 20 10:08:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL081518AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Aug 15 14:58:02 2018
 QLast Update : Wed Aug 15 14:58:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.81	49	1842302	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.35	114	3474133	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.30	117	3485590	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.64	95	2646911	9.95	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.50%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.07	41	228395	3.13	ppbv	93
13) Ethanol	3.70	45	134762	7.18	ppbv	99
15) Acetone	3.95	43	763192	4.36	ppbv	100
19) Isopropyl Alcohol	4.07	45	10236117	86.84	ppbv #	98
26) Hexane	5.83	57	111853	0.67	ppbv #	42
34) 2-Butanone	5.39	43	126937	0.54	ppbv	99
53) Toluene	10.01	91	87028	0.27	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL081618\
Data File : VL032390.D
Acq On : 16 Aug 2018 23:44
Operator : SY/AP
Sample : J4457-08DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
DUP1-MIVIDL

Manual Integrations
APPROVED

MMDadoda
8/20/2018 3:11:53 PM

Quant Time: Aug 20 10:08:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL081518AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Aug
QLast Update : Wed Aug 15 14:58:02 2018
Response via : Initial Calibration

