

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL081618\
 Data File : VL032393.D
 Acq On : 17 Aug 2018 1:43
 Operator : SY/AP
 Sample : J4457-11DL 20X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SS03-MIVIDL

Quant Time: Aug 20 10:17:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL081518AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Aug 15 14:58:02 2018
 QLast Update : Wed Aug 15 14:58:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.81	49	1806628	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.35	114	3453131	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.31	117	3396131	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.64	95	2589968	10.00	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.69	45	3163721	171.98	ppbv	98
15) Acetone	3.95	43	354010	2.06	ppbv	98
19) Isopropyl Alcohol	4.08	45	118081	1.02	ppbv #	92
29) Chloroform	5.90	83	130776	0.54	ppbv	96
34) 2-Butanone	5.39	43	39756	0.17	ppbv	100
35) Carbon Tetrachloride	7.19	117	60224	0.30	ppbv	89
52) Tetrachloroethene	11.44	164	1210166	14.02	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL081618\
Data File : VL032393.D
Acq On : 17 Aug 2018 1:43
Operator : SY/AP
Sample : J4457-11DL 20X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SS03-MIVIDL

Quant Time: Aug 20 10:17:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL081518AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Aug 15 14:58:02 2018
QLast Update : Wed Aug 15 14:58:02 2018
Response via : Initial Calibration

