

Data Path : Y:\VOASRV\HPCHEM1\MSVOA L\DATA\VL081621\
 Data File : VL037466.D
 Acq On : 16 Aug 2021 18:03
 Operator : SY/AP
 Sample : M3277-01
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 E057-1100

Quant Time: Aug 17 11:21:26 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL081621AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Aug 17 02:13:47 2021
 QLast Update : Tue Aug 17 02:13:47 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1355097	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.16	114	4094791	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.07	117	3380876	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.40	95	2663167	11.09	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	110.90%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
4) Chloromethane	3.13	50	2838214	30.608	ppbv	93
11) Trichlorofluoromethane	3.94	101	4119525	20.185	ppbv	97
20) Methylene Chloride	4.34	84	1377844	19.955	ppbv	97
24) 1,1-Dichloroethane	5.00	63	2389607	16.208	ppbv	98
26) Hexane	5.67	57	4483788	26.194	ppbv #	45
35) Carbon Tetrachloride	7.01	117	2594842	10.723	ppbv	98
39) 1,2-Dichloropropane	7.60	63	1846330	13.959	ppbv	94
42) Bromodichloromethane	7.78	83	5746393	23.048	ppbv	95
72) 4-Ethyltoluene	15.82	105	12463743	27.602	ppbv	97
74) 1,2,4-Trimethylbenzene	16.74	105	8156243	18.261	ppbv #	68
77) 1,2-Dichlorobenzene	17.80	146	4118890	15.558	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Y:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL081621\
Data File : VL037466.D
Acq On : 16 Aug 2021 18:03
Operator : SY/AP
Sample : M3277-01
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
E057-1100

Quant Time: Aug 17 11:21:26 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL081621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Aug 17 02:13:47 2021
QLast Update : Tue Aug 17 02:13:47 2021
Response via : Initial Calibration

