


Data Path : Y:\voasrv\HPCHEM1\MSVOA L\Data\VL081621\
 Data File : VL037453.D
 Acq On : 16 Aug 2021 8:41
 Operator : SY/AP
 Sample : BFB
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081621AIR.M
 Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Tue Aug 17 02:13:47 2021
 Last Update : Tue Aug 17 02:13:47 2021



AutoFind: Scans 3608, 3609, 3610; Background Corrected with Scan 3579

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	22.1	221333	PASS
75	95	30	66	53.9	539520	PASS
95	95	100	100	100.0	1000234	PASS
96	95	5	9	7.0	69640	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	70.3	702826	PASS
175	174	4	9	8.4	58965	PASS
176	174	93	101	97.8	687338	PASS
177	176	5	9	6.3	43245	PASS