


Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL081820\
 Data File : VL035443.D
 Acq On : 18 Aug 2020 14:21
 Operator : SY/AP
 Sample : BFB
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081720AIR.M
 Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Mon Aug 17 19:28:30 2020
 Last Update : Mon Aug 17 19:28:30 2020



AutoFind: Scans 3625, 3626, 3627; Background Corrected with Scan 3603

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	14.9	218282	PASS
75	95	30	66	46.7	684394	PASS
95	95	100	100	100.0	1466538	PASS
96	95	5	9	7.5	109610	PASS
173	174	0.00	2	0.6	7639	PASS
174	95	50	120	93.5	1371648	PASS
175	174	4	9	7.5	102957	PASS
176	174	93	101	98.0	1344170	PASS
177	176	5	9	7.2	96349	PASS