

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL081924\
 Data File : VL041540.D
 Acq On : 19 Aug 2024 8:59
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Aug 20 02:13:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Aug 20 02:10:04 2024
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	77	0.00
2 T	Dichlorodifluoromethane	1.720	0.931	45.9#	51	0.00
3	Chlorodifluoromethane	1.190	1.221	-2.6	88	0.00
4	Chloromethane	0.633	0.662	-4.6	91	0.00
5 T	Vinyl Chloride	0.624	0.652	-4.5	89	0.00
6 T	Bromomethane	0.321	0.345	-7.5	91	0.00
7	Chloroethane	0.215	0.241	-12.1	92	0.00
8 T	Dichlorotetrafluoroethane	1.524	1.456	4.5	82	0.00
9 T	Propene	0.547	0.548	-0.2	87	0.00
10 T	Heptane	1.229	1.391	-13.2	90	0.00
11 T	Trichlorofluoromethane	1.528	1.599	-4.6	90	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.100	1.152	-4.7	89	0.00
13	Ethanol	0.049	0.038#	22.4	84	0.00
14 T	Bromoethene	0.429	0.464	-8.2	87	0.00
15 T	Acetone	1.136	1.190	-4.8	92	0.00
16 T	1,3-Butadiene	0.607	0.657	-8.2	92	0.00
17	tert-Butyl alcohol	1.036	1.154	-11.4	85	0.00
18 T	1,1-Dichloroethene	0.471	0.513	-8.9	87	0.00
19 T	Isopropyl Alcohol	0.664	0.686	-3.3	82	0.00
20 T	Methylene Chloride	0.424	0.435	-2.6	91	0.00
21 T	Allyl Chloride	0.745	0.811	-8.9	87	0.00
22 T	trans-1,2-Dichloroethene	0.471	0.517	-9.8	89	0.00
23 T	Vinyl Acetate	0.868	1.017	-17.2	89	0.00
24 T	1,1-Dichloroethane	1.022	1.100	-7.6	91	0.00
25 T	Ethyl Acetate	2.257	2.507	-11.1	92	0.00
26 T	Hexane	0.936	1.031	-10.1	90	0.00
27 T	Carbon Disulfide	1.095	1.224	-11.8	85	0.00
28 T	Methyl tert-Butyl Ether	0.591	0.619	-4.7	86	0.00
29 T	Chloroform	1.527	1.629	-6.7	89	0.00
30 T	Cyclohexane	0.763	0.861	-12.8	87	0.00
31 T	cis-1,2-Dichloroethene	0.925	0.999	-8.0	87	0.00
32 T	1,1,1-Trichloroethane	1.487	1.607	-8.1	89	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	76	0.00
34 T	2-Butanone	0.484	0.549	-13.4	90	0.00
35 T	Carbon Tetrachloride	0.552	0.629	-13.9	88	0.00
36 T	Benzene	0.731	0.805	-10.1	87	0.00
37 T	1,2-Dichloroethane	0.409	0.451	-10.3	88	0.00
38 T	Trichloroethene	0.335	0.382	-14.0	91	0.00
39 T	1,2-Dichloropropane	0.280	0.317	-13.2	91	0.00
40 T	1,4-Dioxane	0.112	0.122	-8.9	81	0.00
41 T	Tetrahydrofuran	0.277	0.325	-17.3	89	0.00
42 T	Bromodichloromethane	0.591	0.662	-12.0	88	0.00
43	Methyl Methacrylate	0.262	0.321	-22.5	87	0.00
44 T	2,2,4-Trimethylpentane	1.220	1.373	-12.5	90	0.00
45 T	t-1,3-Dichloropropene	0.241	0.329	-36.5#	88	0.00
46 T	cis-1,3-Dichloropropene	0.348	0.435	-25.0	87	0.00
47 T	1,1,2-Trichloroethane	0.308	0.342	-11.0	89	0.00
48 T	Dibromochloromethane	0.497	0.575	-15.7	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL081924\
 Data File : VL041540.D
 Acq On : 19 Aug 2024 8:59
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Aug 20 02:13:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Aug 20 02:10:04 2024
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.420	0.504	-20.0	86	0.00
50 T	4-Methyl-2-Pentanone	0.763	0.917	-20.2	90	0.00
51 T	2-Hexanone	0.577	0.738	-27.9	88	0.00
52 T	Tetrachloroethene	0.282	0.307	-8.9	86	0.00
53 T	Toluene	0.814	0.951	-16.8	88	0.00
54 T	1,2-Dibromoethane	0.410	0.500	-22.0	89	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	77	0.00
56	1,1,1,2-Tetrachloroethane	0.388	0.430	-10.8	89	0.00
57 T	Chlorobenzene	0.712	0.762	-7.0	88	0.00
58 T	Ethyl Benzene	1.153	1.310	-13.6	88	0.00
59 T	m/p-Xylene	0.995	1.128	-13.4	89	0.00
60 T	o-Xylene	0.941	1.080	-14.8	89	0.00
61 T	Styrene	0.425	0.522	-22.8	87	0.00
62	Isopropylbenzene	1.431	1.568	-9.6	88	0.00
63 T	1,1,2,2-Tetrachloroethane	0.594	0.666	-12.1	90	0.00
64	n-propylbenzene	0.374	0.420	-12.3	88	0.00
65	tert-Butylbenzene	1.292	1.419	-9.8	89	0.00
66 T	Benzyl Chloride	0.083	0.117	-41.0#	98	0.00
67	sec-Butylbenzene	1.781	1.992	-11.8	89	0.00
68 S	1-Bromo-4-Fluorobenzene	0.770	0.774	-0.5	76	0.00
69	p-Isopropyltoluene	1.478	1.679	-13.6	89	0.00
70	n-Butylbenzene	1.515	1.723	-13.7	89	0.00
71	2-Chlorotoluene	1.055	1.186	-12.4	89	0.00
72 T	4-Ethyltoluene	1.108	1.284	-15.9	87	0.00
73 T	1,3,5-Trimethylbenzene	0.955	1.111	-16.3	88	0.00
74 T	1,2,4-Trimethylbenzene	1.110	1.247	-12.3	90	0.00
75 T	1,3-Dichlorobenzene	0.698	0.777	-11.3	89	0.00
76 T	1,4-Dichlorobenzene	0.676	0.769	-13.8	90	0.00
77 T	1,2-Dichlorobenzene	0.692	0.755	-9.1	89	0.00
78 T	Hexachloro-1,3-Butadiene	0.590	0.563	4.6	87	0.00
79 T	Naphthalene	0.980	1.250	-27.6	86	0.00
80 T	Naphthalene,2-methyl-	0.821	1.072	-30.6#	83	0.00
81 T	1,2,4-Trichlorobenzene	0.583	0.661	-13.4	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0