

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL083118\
 Data File : VL032444.D
 Acq On : 31 Aug 2018 17:02
 Operator : SY/AP
 Sample : J4472-02DL 10X
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SS-2DL

Manual Integrations
 APPROVED

apatel
 9/4/2018 8:26:24 AM

Quant Time: Sep 02 22:14:18 2018

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL083118AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 31 2018

QLast Update : Fri Aug 31 21:33:46 2018

Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.80	49	2026678	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.34	114	3878599	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.29	117	3712358	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.63	95	2836271	10.29	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.90%

Target Compounds

						Qvalue
9) Propene	3.07	41	44240	0.598	ppbv	99
11) Trichlorofluoromethane	4.04	101	68411	0.293	ppbv	93
13) Ethanol	3.69	45	79123	3.430	ppbv	96
15) Acetone	3.94	43	338724	1.856	ppbv	100
19) Isopropyl Alcohol	4.08	45	185537	1.550	ppbv #	95
34) 2-Butanone	5.38	43	93536	0.363	ppbv	99
53) Toluene	9.99	91	47783m	0.138	ppbv	
76) 1,4-Dichlorobenzene	17.36	146	855233	3.339	ppbv	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL083118\
Data File : VL032444.D
Acq On : 31 Aug 2018 17:02
Operator : SY/AP
Sample : J4472-02DL 10X
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SS-2DL

Manual Integrations
APPROVED

apatel
9/4/2018 8:26:24 AM

Quant Time: Sep 02 22:14:18 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL083118AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Aug 31 21:33:46 2018
Response via : Initial Calibration

