

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL090320\
 Data File : VL035530.D
 Acq On : 3 Sep 2020 14:30
 Operator : SY/AP
 Sample : OC090220 CAN 10401
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC090220 CAN 10401

Manual Integrations
 APPROVED

MMDadoda
 9/8/2020 4:06:12 PM

Quant Time: Sep 04 07:04:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL082720AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Aug 27 2020
 QLast Update : Thu Aug 27 06:46:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.68	49	1041882	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.20	114	4416771	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.12	117	4437560	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.45	95	3226345	10.24	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.40%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.63	45	4629m	0.420	ppbv	
15) Acetone	3.88	43	44587	0.388	ppbv	# 49
20) Methylene Chloride	4.36	84	67144	0.785	ppbv	98
26) Hexane	5.71	57	12609m	0.110	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL090320\
Data File : VL035530.D
Acq On : 3 Sep 2020 14:30
Operator : SY/AP
Sample : OC090220 CAN 10401
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC090220 CAN 10401

Manual Integrations
APPROVED

MMDadoda
9/8/2020 4:06:12 PM

Quant Time: Sep 04 07:04:48 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL082720AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Aug
QLast Update : Thu Aug 27 06:46:28 2020
Response via : Initial Calibration

