

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL090321\
Data File : VL037519.D
Acq On : 3 Sep 2021 13:50
Operator : SY/AP
Sample : CAN 10053
Misc : 400mL/MSVOA_L
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CAN 10053

Quant Time: Sep 04 04:21:59 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LTue Aug 17 02:13:47 2021
QLast Update : Tue Aug 17 02:13:47 2021
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.661	49	1454110	10.000	ppbv	0.01
33) 1,4-Difluorobenzene	7.169	114	4209887	10.000	ppbv	0.01
55) Chlorobenzene-d5	12.076	117	3494434	10.000	ppbv	0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.407	95	2495306	10.051	ppbv	0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL090321\
Data File : VL037519.D
Acq On : 3 Sep 2021 13:50
Operator : SY/AP
Sample : CAN 10053
Misc : 400mL/MSVOA_L
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CAN 10053

Quant Time: Sep 04 04:21:59 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Aug 17 02:13:47 2021
QLast Update : Tue Aug 17 02:13:47 2021
Response via : Initial Calibration

