

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL090321\
 Data File : VL037524.D
 Acq On : 3 Sep 2021 16:57
 Operator : SY/AP
 Sample : CAN 10596
 Misc : 400mL/MSVOA_L
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 CAN 10596

Quant Time: Sep 04 04:22:30 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081621AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Tue Aug 17 02:13:47 2021
 QLast Update : Tue Aug 17 02:13:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.665	49	1579300	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	7.169	114	4865236	10.000	ppbv	0.01
55) Chlorobenzene-d5	12.079	117	4235095	10.000	ppbv	0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.407	95	3117701	10.362	ppbv	0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL090321\
Data File : VL037524.D
Acq On : 3 Sep 2021 16:57
Operator : SY/AP
Sample : CAN 10596
Misc : 400mL/MSVOA_L
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CAN 10596

Quant Time: Sep 04 04:22:30 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Aug 17 02:13:47 2021
QLast Update : Tue Aug 17 02:13:47 2021
Response via : Initial Calibration

