

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL090321\
Data File : VL037528.D
Acq On : 3 Sep 2021 19:27
Operator : SY/AP
Sample : CAN 10300
Misc : 400mL/MSVOA_L
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CAN 10300

Quant Time: Sep 04 04:22:56 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Aug 17 02:13:47 2021
QLast Update : Tue Aug 17 02:13:47 2021
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.661	49	1657256	10.000	ppbv	0.01
33) 1,4-Difluorobenzene	7.169	114	4770144	10.000	ppbv	0.01
55) Chlorobenzene-d5	12.079	117	4085100	10.000	ppbv	0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.407	95	3041915	10.481	ppbv	0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	104.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL090321\
Data File : VL037528.D
Acq On : 3 Sep 2021 19:27
Operator : SY/AP
Sample : CAN 10300
Misc : 400mL/MSVOA_L
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CAN 10300

Quant Time: Sep 04 04:22:56 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Aug 17 02:13:47 2021
QLast Update : Tue Aug 17 02:13:47 2021
Response via : Initial Calibration

