

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL090321\
Data File : VL037532.D
Acq On : 3 Sep 2021 22:00
Operator : SY/AP
Sample : CAN 10599
Misc : 400mL/MSVOA_L
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CAN 10599

Quant Time: Sep 04 04:23:25 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Aug 17 02:13:47 2021
QLast Update : Tue Aug 17 02:13:47 2021
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.661	49	1599145	10.000	ppbv	0.01
33) 1,4-Difluorobenzene	7.169	114	4735038	10.000	ppbv	0.01
55) Chlorobenzene-d5	12.076	117	4135648	10.000	ppbv	0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.404	95	3065141	10.432	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	104.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL090321\
Data File : VL037532.D
Acq On : 3 Sep 2021 22:00
Operator : SY/AP
Sample : CAN 10599
Misc : 400mL/MSVOA_L
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CAN 10599

Quant Time: Sep 04 04:23:25 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Aug 17 02:13:47 2021
QLast Update : Tue Aug 17 02:13:47 2021
Response via : Initial Calibration

