

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL090321\
 Data File : VL037534.D
 Acq On : 3 Sep 2021 23:17
 Operator : SY/AP
 Sample : CAN 10608
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 CAN 10608

Quant Time: Sep 04 04:23:51 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081621AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Aug 17 02:13:47 2021
 QLast Update : Tue Aug 17 02:13:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.661	49	1576600	10.000	ppbv	0.01
33) 1,4-Difluorobenzene	7.166	114	4670344	10.000	ppbv	0.00
55) Chlorobenzene-d5	12.076	117	4040229	10.000	ppbv	0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.407	95	3036261	10.578	ppbv	0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	105.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL090321\
Data File : VL037534.D
Acq On : 3 Sep 2021 23:17
Operator : SY/AP
Sample : CAN 10608
Misc : 400mL/MSVOA_L
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CAN 10608

Quant Time: Sep 04 04:23:51 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081621AIR.M
Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Tue Aug 17 02:13:47 2021
QLast Update : Tue Aug 17 02:13:47 2021
Response via : Initial Calibration

