

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL090425\
 Data File : VL042895.D
 Acq On : 04 Sep 2025 18:05
 Operator : SY/MD
 Sample : Q3000-03DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA2DL

Quant Time: Sep 05 01:24:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL081325AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Thu Aug 14 02:12:54 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.800	49	121196	10.000	ppbv	0.01
33) 1,4-Difluorobenzene	3.978	114	316343	10.000	ppbv	0.01
55) Chlorobenzene-d5	8.901	117	265622	10.000	ppbv	0.01
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.393	95	199318	10.005	ppbv	0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.100%
Target Compounds						
					Qvalue	
4) Chloromethane	1.538	50	872	0.122	ppbv #	84
13) Ethanol	1.729	45	16239	20.946	ppbv	95
15) Acetone	1.845	43	11606	0.724	ppbv	98
20) Methylene Chloride	2.072	84	1893	0.282	ppbv #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL090425\
Data File : VL042895.D
Acq On : 04 Sep 2025 18:05
Operator : SY/MD
Sample : Q3000-03DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA2DL

Quant Time: Sep 05 01:24:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL081325AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Thu Aug 14 02:12:54 2025
Response via : Initial Calibration

