

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL091018\
 Data File : VL032501.D
 Acq On : 11 Sep 2018 10:26
 Operator : SY/AP
 Sample : J4820-05DL2 100X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 249-BERRYDL2

Manual Integrations
 APPROVED

MMDadoda
 9/12/2018 3:59:52 PM

Quant Time: Sep 12 07:23:46 2018

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL083118AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 31 21:33:46 2018

QLast Update : Fri Aug 31 21:33:46 2018

Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.80	49	1356581	10.00	ppbv	-0.01
33) 1,4-Difluorobenzene	7.33	114	3129023	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.28	117	3151193	10.00	ppbv	-0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.62	95	2456477	10.50	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	105.00%

Target Compounds

	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.06	41	117653	2.38	ppbv	92
13) Ethanol	3.69	45	17083m	1.11	ppbv	
15) Acetone	3.93	43	290819	2.38	ppbv #	72
29) Chloroform	5.89	83	36667	0.19	ppbv	84
34) 2-Butanone	5.36	43	2063994	9.93	ppbv	100
51) 2-Hexanone	10.33	43	161208m	0.80	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL091018\
Data File : VL032501.D
Acq On : 11 Sep 2018 10:26
Operator : SY/AP
Sample : J4820-05DL2 100X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
249-BERRYDL2

Manual Integrations
APPROVED

MMDadoda
9/12/2018 3:59:52 PM

Quant Time: Sep 12 07:23:46 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL083118AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Aug 31 21:33:46 2018
Response via : Initial Calibration

