

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL091420\
 Data File : VL035611.D
 Acq On : 14 Sep 2020 22:55
 Operator : SY/AP
 Sample : VIBLK
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Manual Integrations
 APPROVED

MMDadoda
 9/17/2020 5:40:44 PM

Quant Time: Sep 15 07:45:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL082720AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Sep 14 2020
 QLast Update : Mon Sep 14 15:19:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1045667	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.16	114	4497853	10.00	ppbv	-0.01
55) Chlorobenzene-d5	12.08	117	4494765	10.00	ppbv	-0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.41	95	3294644	10.32	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.20%
Target Compounds						
20) Methylene Chloride	4.33	84	14871	0.173	ppbv	96
52) Tetrachloroethene	11.22	164	8637m	0.051	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL091420\
Data File : VL035611.D
Acq On : 14 Sep 2020 22:55
Operator : SY/AP
Sample : VIBLK
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Manual Integrations
APPROVED

MMDadoda
9/17/2020 5:40:44 PM

Quant Time: Sep 15 07:45:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL082720AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Sep
QLast Update : Mon Sep 14 15:19:50 2020
Response via : Initial Calibration

