

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL091420\
 Data File : VL035617.D
 Acq On : 15 Sep 2020 2:41
 Operator : SY/AP
 Sample : VIBLK
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Manual Integrations
 APPROVED

MMDadoda
 9/17/2020 5:41:03 PM

Quant Time: Sep 15 08:06:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL082720AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Sep 14 15:19:50 2020
 QLast Update : Mon Sep 14 15:19:50 2020
 Response via : Initial Calibration

Internal Standards			R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane			5.65	49	990092	10.00	ppbv	-0.01
33) 1,4-Difluorobenzene			7.16	114	4309561	10.00	ppbv	-0.01
55) Chlorobenzene-d5			12.08	117	4354643	10.00	ppbv	-0.02
System Monitoring Compounds								
68) 1-Bromo-4-Fluorobenzene			14.41	95	3190702	10.32	ppbv	-0.02
Spiked Amount		10.000	Range	65 - 135	Recovery	=	103.20%	
Target Compounds								
20) Methvlene Chloride			4.33	84	14986m	0.184	ppbv	Ovalue
38) Trichloroethene			7.82	130	6171	0.038	ppbv	# 80
52) Tetrachloroethene			11.21	164	45684	0.280	ppbv	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL091420\
Data File : VL035617.D
Acq On : 15 Sep 2020 2:41
Operator : SY/AP
Sample : VIBLK
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Manual Integrations
APPROVED

MMDadoda
9/17/2020 5:41:03 PM

Quant Time: Sep 15 08:06:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL082720AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Sep
QLast Update : Mon Sep 14 15:19:50 2020
Response via : Initial Calibration

