


Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL091420\
 Data File : VL035595.D
 Acq On : 14 Sep 2020 9:30
 Operator : SY/AP
 Sample : BFB
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL082720AIR.M
 Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Mon Sep 14 15:19:50 2020
 Last Update : Mon Sep 14 15:19:50 2020



AutoFind: Scans 3618, 3619, 3620; Background Corrected with Scan 3594

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	15.7	227498	PASS
75	95	30	66	48.4	702933	PASS
95	95	100	100	100.0	1452544	PASS
96	95	5	9	7.5	109101	PASS
173	174	0.00	2	0.6	7051	PASS
174	95	50	120	79.9	1160533	PASS
175	174	4	9	7.6	88762	PASS
176	174	93	101	94.9	1101141	PASS
177	176	5	9	7.0	76962	PASS