

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL091620\
 Data File : VL035638.D
 Acq On : 16 Sep 2020 21:04
 Operator : SY/AP
 Sample : L4036-01DL 4X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 DB-IA-Y01DL

Manual Integrations
 APPROVED

MMDadoda
 9/17/2020 1:52:43 PM

Quant Time: Sep 17 07:55:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL082720AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Sep 14 15:19:50 2020
 QLast Update : Mon Sep 14 15:19:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1262803	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.17	114	4930734	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.09	117	4881916	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.42	95	3588948	10.35	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.50%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
4) Chloromethane	3.13	50	11158	0.149	ppbv	96
9) Propene	3.02	41	6914m	0.128	ppbv	
13) Ethanol	3.60	45	39112	2.930	ppbv	96
15) Acetone	3.85	43	122335	0.878	ppbv	89
19) Isopropyl Alcohol	3.98	45	28973	0.349	ppbv #	84
20) Methylene Chloride	4.34	84	82589	0.796	ppbv	93
26) Hexane	5.68	57	25282m	0.181	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL091620\
Data File : VL035638.D
Acq On : 16 Sep 2020 21:04
Operator : SY/AP
Sample : L4036-01DL 4X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
DB-IA-Y01DL

Manual Integrations
APPROVED

MMDadoda
9/17/2020 1:52:43 PM

Quant Time: Sep 17 07:55:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL082720AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Sep
QLast Update : Mon Sep 14 15:19:50 2020
Response via : Initial Calibration

