

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL091620\
 Data File : VL035644.D
 Acq On : 17 Sep 2020 1:39
 Operator : SY/AP
 Sample : VIBLK
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Quant Time: Sep 17 08:09:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL082720AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Sep 14 15:19:50 2020
 QLast Update : Mon Sep 14 15:19:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1152894	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.17	114	4429348	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.09	117	4478407	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.43	95	3269609	10.28	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.80%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	3.87	43	21613	0.170	ppbv	# 44
20) Methylene Chloride	4.34	84	17950	0.190	ppbv	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL091620\
Data File : VL035644.D
Acq On : 17 Sep 2020 1:39
Operator : SY/AP
Sample : VIBLK
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: Sep 17 08:09:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL082720AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Sep 14 15:19:50 2020
QLast Update : Mon Sep 14 15:19:50 2020
Response via : Initial Calibration

