

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL091720\
 Data File : VL035656.D
 Acq On : 17 Sep 2020 13:48
 Operator : SY/AP
 Sample : VL0917ABL01
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0917ABL01

Quant Time: Sep 18 04:09:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL082720AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Sep 14 15:19:50 2020
 QLast Update : Mon Sep 14 15:19:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1116015	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	4380396	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.10	117	4373467	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.43	95	3176201	10.23	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.30%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	3.87	43	30863	0.251	ppbv	# 32
20) Methylene Chloride	4.35	84	16575	0.181	ppbv	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL091720\
Data File : VL035656.D
Acq On : 17 Sep 2020 13:48
Operator : SY/AP
Sample : VL0917ABL01
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0917ABL01

Quant Time: Sep 18 04:09:36 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL082720AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Sep 14 15:19:50 2020
QLast Update : Mon Sep 14 15:19:50 2020
Response via : Initial Calibration

