

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL091721\
 Data File : VL037646.D
 Acq On : 17 Sep 2021 18:01
 Operator : SY/AP
 Sample : M3801-02DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA1DL

Manual Integrations
 APPROVED

MMDadoda
 9/20/2021 5:28:41 PM

Quant Time: Sep 18 07:39:49 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL091621AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Sep 1
 QLast Update : Thu Sep 16 13:42:46 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1386503	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.16	114	3999212	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.07	117	3370926	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.40	95	2363295	9.70	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.60	45	394820m	43.755	ppbv	
15) Acetone	3.85	43	201931m	1.402	ppbv	
19) Isopropyl Alcohol	3.96	45	823722	10.509	ppbv	98
20) Methylene Chloride	4.33	84	273147	3.972	ppbv	98
26) Hexane	5.67	57	83313	0.503	ppbv #	52

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL091721\
Data File : VL037646.D
Acq On : 17 Sep 2021 18:01
Operator : SY/AP
Sample : M3801-02DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA1DL

Manual Integrations
APPROVED

MMDadoda
9/20/2021 5:28:41 PM

Quant Time: Sep 18 07:39:49 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL091621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Sep
QLast Update : Thu Sep 16 13:42:46 2021
Response via : Initial Calibration

