

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL091721\
 Data File : VL037648.D
 Acq On : 17 Sep 2021 19:20
 Operator : SY/AP
 Sample : M3801-03DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 OA1DL

Manual Integrations
 APPROVED

MMDadoda
 9/20/2021 5:28:45 PM

Quant Time: Sep 18 07:49:09 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL091621AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Sep 16 13:42:46 2021
 QLast Update : Thu Sep 16 13:42:46 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1296952	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.16	114	3780905	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.07	117	3297771	10.00	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.40	95	2231492	9.36	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	93.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.61	45	34097m	4.040	ppbv	
15) Acetone	3.86	43	92688m	0.688	ppbv	
19) Isopropyl Alcohol	3.98	45	37449m	0.511	ppbv	
20) Methylene Chloride	4.34	84	58689m	0.912	ppbv	
26) Hexane	5.68	57	28681	0.185	ppbv #	52

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL091721\
Data File : VL037648.D
Acq On : 17 Sep 2021 19:20
Operator : SY/AP
Sample : M3801-03DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
OA1DL

Manual Integrations
APPROVED

MMDadoda
9/20/2021 5:28:45 PM

Quant Time: Sep 18 07:49:09 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL091621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Sep
QLast Update : Thu Sep 16 13:42:46 2021
Response via : Initial Calibration

