

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL091721\
 Data File : VL037650.D
 Acq On : 18 Sep 2021 8:46
 Operator : SY/AP
 Sample : OC091321 CAN 10428
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC091321 CAN 10428

Manual Integrations
 APPROVED

MMDadoda
 9/20/2021 5:28:49 PM

Quant Time: Sep 20 11:26:00 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL091621AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Thu Sep 16 2021
 QLast Update : Thu Sep 16 13:42:46 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1305541	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.17	114	3817400	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.08	117	3304054	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.40	95	2276235	9.53	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	95.30%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL091721\
Data File : VL037650.D
Acq On : 18 Sep 2021 8:46
Operator : SY/AP
Sample : QC091321 CAN 10428
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC091321 CAN 10428

Manual Integrations
APPROVED

MMDadoda
9/20/2021 5:28:49 PM

Quant Time: Sep 20 11:26:00 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL091621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Sep
QLast Update : Thu Sep 16 13:42:46 2021
Response via : Initial Calibration

