


Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL091721\
 Data File : VL037637.D
 Acq On : 17 Sep 2021 11:30
 Operator : SY/AP
 Sample : BFB
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL091621AIR.M
 Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Thu Sep 16 13:42:46 2021
 Last Update : Thu Sep 16 13:42:46 2021



AutoFind: Scans 3608, 3609, 3610; Background Corrected with Scan 3585

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	22.3	202304	PASS
75	95	30	66	53.0	480853	PASS
95	95	100	100	100.0	906944	PASS
96	95	5	9	6.7	61026	PASS
173	174	0.00	2	0.8	4916	PASS
174	95	50	120	69.7	632554	PASS
175	174	4	9	8.0	50597	PASS
176	174	93	101	95.3	602752	PASS
177	176	5	9	6.3	38042	PASS