

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091725\
 Data File : VL042935.D
 Acq On : 17 Sep 2025 13:42
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/18/2025
 Supervised By :Mahesh Dadoda 09/18/2025

Quant Time: Sep 18 02:32:00 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091725AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2025

QLast Update : Thu Sep 18 02:25:28 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.790	49	160573	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.965	114	415318	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.888	117	353249	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.380 95 295803 10.713 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 107.100%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.502	85	232093	10.119	ppbv	100
3) Chlorodifluoromethane	1.479	51	238232	9.865	ppbv	100
4) Chloromethane	1.534	50	92202	10.003	ppbv	100
5) Vinyl Chloride	1.583	62	80228	9.226	ppbv	100
6) Bromomethane	1.670	94	30537	10.105	ppbv	100
7) Chloroethane	1.706	64	32281	10.079	ppbv	100
8) Dichlorotetrafluoroethane	1.557	85	183602	10.089	ppbv	100
9) Propene	1.486	41	86426	9.800	ppbv	100
10) Heptane	4.988	43	259422	11.872	ppbv	100
11) Trichlorofluoromethane	1.881	101	229480	10.128	ppbv	100
12) 1,1,2-Trichlorotrifluo...	2.143	101	178790	10.100	ppbv	100
13) Ethanol	1.722	45	7630	8.486	ppbv	98
14) Bromoethene	1.784	108	62744	9.949	ppbv	100
15) Acetone	1.839	43	172060	8.724	ppbv	100
16) 1,3-Butadiene	1.612	39	79400	9.624	ppbv	100
17) tert-Butyl alcohol	2.042	59	187824	9.890	ppbv	100
18) 1,1-Dichloroethene	2.036	96	78341	10.110	ppbv	100
19) Isopropyl Alcohol	1.887	45	103481	9.786	ppbv	100
20) Methylene Chloride	2.065	84	67680	7.882	ppbv	100
21) Allyl Chloride	2.097	41	132660	9.923	ppbv	100
22) trans-1,2-Dichloroethene	2.343	96	88151	10.037	ppbv	100
23) Vinyl Acetate	2.466	43	238255	10.397	ppbv	100
24) 1,1-Dichloroethane	2.411	63	176500	10.174	ppbv	100
25) Ethyl Acetate	2.835	43	455318	10.811	ppbv	100
26) Hexane	2.829	57	199382	11.031	ppbv	100
27) Carbon Disulfide	2.152	76	220787	9.904	ppbv	99
28) Methyl tert-Butyl Ether	2.444	73	99133	9.982	ppbv	100
29) Chloroform	2.852	83	299464	10.137	ppbv	100
30) Cyclohexane	3.861	84	162355	11.924	ppbv	100
31) cis-1,2-Dichloroethene	2.725	61	184566	10.459	ppbv	100
32) 1,1,1-Trichloroethane	3.373	97	274368	9.926	ppbv	100
34) 2-Butanone	2.560	43	280052	10.402	ppbv	100
35) Carbon Tetrachloride	3.768	117	287073	9.724	ppbv	100
36) Benzene	3.661	78	383581	10.588	ppbv	100
37) 1,2-Dichloroethane	3.221	62	218944	10.091	ppbv	100
38) Trichloroethene	4.557	130	166868	10.176	ppbv	100
39) 1,2-Dichloropropane	4.318	63	148415	10.139	ppbv	100
40) 1,4-Dioxane	4.583	88	60929	11.733	ppbv #	100
41) Tetrahydrofuran	3.056	42	150567	11.375	ppbv	100
42) Bromodichloromethane	4.496	83	316501	10.165	ppbv	100
43) Methyl Methacrylate	4.884	69	143457	11.807	ppbv	100
44) 2,2,4-Trimethylpentane	4.648	57	700623	11.595	ppbv	100
45) t-1,3-Dichloropropene	6.422	75	146102	11.424	ppbv	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091725\
 Data File : VL042935.D
 Acq On : 17 Sep 2025 13:42
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/18/2025
 Supervised By :Mahesh Dadoda 09/18/2025

Quant Time: Sep 18 02:32:00 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091725AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2025

QLast Update : Thu Sep 18 02:25:28 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.609	75	197003	11.174	ppbv	100
47) 1,1,2-Trichloroethane	6.587	97	156042	10.453	ppbv	100
48) Dibromochloromethane	7.393	129	244251	10.447	ppbv	100
49) Bromoform	9.490	173	196118	10.747	ppbv	100
50) 4-Methyl-2-Pentanone	5.758	43	387522	12.080	ppbv	100
51) 2-Hexanone	7.441	43	307420	13.020	ppbv	100
52) Tetrachloroethene	8.231	164	115985	10.295	ppbv	100
53) Toluene	6.939	91	415754	12.313	ppbv	100
54) 1,2-Dibromoethane	7.658	107	211125	10.997	ppbv	100
56) 1,1,1,2-Tetrachloroethane	8.930	131	186222	10.345	ppbv	100
57) Chlorobenzene	8.927	112	342500	10.490	ppbv	100
58) Ethyl Benzene	9.360	91	582982	12.810	ppbv	100
59) m/p-Xylene	9.558	91	983541m	25.571	ppbv	
60) o-Xylene	9.969	91	495406	12.957	ppbv	100
61) Styrene	9.875	104	196509	9.941	ppbv	100
62) Isopropylbenzene	10.545	105	707651	12.705	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.966	83	283680	10.272	ppbv	100
64) n-propylbenzene	10.992	120	182766	12.656	ppbv	100
65) tert-Butylbenzene	11.536	119	619284	13.011	ppbv	100
66) Benzyl Chloride	11.620	91	53013	11.987	ppbv	100
67) sec-Butylbenzene	11.772	105	904817	12.496	ppbv	100
69) p-Isopropyltoluene	11.930	119	730122	13.084	ppbv	100
70) n-Butylbenzene	12.286	91	787146	10.537	ppbv	100
71) 2-Chlorotoluene	10.904	91	530980	12.848	ppbv	100
72) 4-Ethyltoluene	11.131	105	597492	10.333	ppbv	100
73) 1,3,5-Trimethylbenzene	11.209	105	493710	13.051	ppbv	100
74) 1,2,4-Trimethylbenzene	11.539	105	566106	12.734	ppbv	100
75) 1,3-Dichlorobenzene	11.613	146	319203	11.326	ppbv	100
76) 1,4-Dichlorobenzene	11.678	146	322554	11.896	ppbv	100
77) 1,2-Dichlorobenzene	11.950	146	316961	11.147	ppbv	100
78) Hexachloro-1,3-Butadiene	13.885	225	185221	10.199	ppbv	100
79) Naphthalene	13.516	128	474749	10.151	ppbv	100
80) Naphthalene,2-methyl-	14.461	142	132693	9.058	ppbv	100
81) 1,2,4-Trichlorobenzene	13.445	180	224163	9.993	ppbv	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

