

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL092120\
 Data File : VL035686.D
 Acq On : 21 Sep 2020 15:12
 Operator : SY/AP
 Sample : L4016-01DL2DUP 400X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV-13DL2DUP

Quant Time: Sep 22 11:45:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL082720AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Sep 14 15:19:50 2020
 QLast Update : Mon Sep 14 15:19:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.63	49	1033098	10.00	ppbv	-0.03
33) 1,4-Difluorobenzene	7.14	114	3939168	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.06	117	3966220	10.00	ppbv	-0.04

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.39	95	2894916	10.28	ppbv	-0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.80%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	3.83	43	104296	0.915	ppbv	92
38) Trichloroethene	7.80	130	31235	0.211	ppbv	97
52) Tetrachloroethene	11.20	164	1949315	13.086	ppbv	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL092120\
Data File : VL035686.D
Acq On : 21 Sep 2020 15:12
Operator : SY/AP
Sample : L4016-01DL2DUP 400X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV-13DL2DUP

Quant Time: Sep 22 11:45:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL082720AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Sep 14 15:19:50 2020
QLast Update : Mon Sep 14 15:19:50 2020
Response via : Initial Calibration

