

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL092120\
 Data File : VL035687.D
 Acq On : 21 Sep 2020 15:50
 Operator : SY/AP
 Sample : VIBLK
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Quant Time: Sep 22 01:53:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL082720AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Sep 14 15:19:50 2020
 QLast Update : Mon Sep 14 15:19:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.64	49	1034121	10.00	ppbv	-0.03
33) 1,4-Difluorobenzene	7.14	114	3927284	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.06	117	3882987	10.00	ppbv	-0.04

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.39	95	2882801	10.45	ppbv	-0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	104.50%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	3.85	43	14962	0.131	ppbv	# 44
20) Methylene Chloride	4.31	84	17582	0.207	ppbv	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL092120\
Data File : VL035687.D
Acq On : 21 Sep 2020 15:50
Operator : SY/AP
Sample : VIBLK
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: Sep 22 01:53:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL082720AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Sep 14 15:19:50 2020
QLast Update : Mon Sep 14 15:19:50 2020
Response via : Initial Calibration

