

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL092120\
 Data File : VL035699.D
 Acq On : 21 Sep 2020 23:25
 Operator : SY/AP
 Sample : L3983-01DL2 1200X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV-8DL2

Quant Time: Sep 22 10:51:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL082720AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Sep 14 15:19:50 2020
 QLast Update : Mon Sep 14 15:19:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.64	49	997645	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.15	114	3624012	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.07	117	3697452	10.00	ppbv	-0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.40	95	2671669	10.17	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.70%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
31) cis-1,2-Dichloroethene	5.53	61	340486	3.013	ppbv	96
38) Trichloroethene	7.81	130	1878093	13.783	ppbv	93
52) Tetrachloroethene	11.22	164	7973116	58.179	ppbv #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL092120\
Data File : VL035699.D
Acq On : 21 Sep 2020 23:25
Operator : SY/AP
Sample : L3983-01DL2 1200X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV-8DL2

Quant Time: Sep 22 10:51:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL082720AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Sep 14 15:19:50 2020
QLast Update : Mon Sep 14 15:19:50 2020
Response via : Initial Calibration

