


Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL092120\
 Data File : VL035681.D
 Acq On : 21 Sep 2020 9:57
 Operator : SY/AP
 Sample : BFB
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL082720AIR.M
 Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Tue Sep 22 01:39:30 2020
 Last Update : Tue Sep 22 01:39:30 2020



AutoFind: Scans 3606, 3607, 3608; Background Corrected with Scan 3582

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	16.3	195520	PASS
75	95	30	66	48.4	580053	PASS
95	95	100	100	100.0	1199616	PASS
96	95	5	9	7.2	86008	PASS
173	174	0.00	2	0.5	4488	PASS
174	95	50	120	76.3	915541	PASS
175	174	4	9	7.6	69128	PASS
176	174	93	101	96.6	884693	PASS
177	176	5	9	7.0	61658	PASS