

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092225\
 Data File : VL042965.D
 Acq On : 22 Sep 2025 19:49
 Operator : SY/MD
 Sample : Q3112-02DL 40X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV2DL

Quant Time: Sep 23 02:54:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Sep 23 02:39:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.790	49	159259	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.965	114	479324	10.000	ppbv	0.01
55) Chlorobenzene-d5	8.885	117	422943	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.380	95	307794	9.749	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.500%
Target Compounds						
						Qvalue
11) Trichlorofluoromethane	1.881	101	2884	0.140	ppbv	99
15) Acetone	1.839	43	16126	0.777	ppbv	97
20) Methylene Chloride	2.065	84	20063	2.759	ppbv	96
26) Hexane	2.829	57	25506	1.067	ppbv #	40

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092225\
Data File : VL042965.D
Acq On : 22 Sep 2025 19:49
Operator : SY/MD
Sample : Q3112-02DL 40X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV2DL

Quant Time: Sep 23 02:54:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Tue Sep 23 02:39:16 2025
Response via : Initial Calibration

