

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092225\
 Data File : VL042981.D
 Acq On : 23 Sep 2025 05:19
 Operator : SY/MD
 Sample : Q3132-02DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 151-1A-2DL

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 09/24/2025
 Supervised By :Semsettin Yesilyurt 09/24/2025

Quant Time: Sep 24 01:17:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Tue Sep 23 02:39:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.790	49	142477	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.965	114	408754	10.000	ppbv	0.01
55) Chlorobenzene-d5	8.885	117	352010	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.380	95	267858	10.194	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.900%
Target Compounds						
					Qvalue	
13) Ethanol	1.722	45	35047	34.770	ppbv	93
15) Acetone	1.839	43	25846m	1.393	ppbv	
19) Isopropyl Alcohol	1.891	45	88410	7.925	ppbv #	39
26) Hexane	2.832	57	6872	0.321	ppbv #	55
53) Toluene	6.940	91	5680m	0.125	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092225\
Data File : VL042981.D
Acq On : 23 Sep 2025 05:19
Operator : SY/MD
Sample : Q3132-02DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
151-1A-2DL

Quant Time: Sep 24 01:17:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Tue Sep 23 02:39:16 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 09/24/2025
Supervised By :Semsettin Yesilyurt 09/24/2025

