

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL092321\
 Data File : VL037655.D
 Acq On : 23 Sep 2021 10:22
 Operator : SY/AP
 Sample : VL0923ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0923ABL01

Quant Time: Sep 24 06:45:21 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL091621AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Thu Sep 16 13:42:46 2021
 QLast Update : Thu Sep 16 13:42:46 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1600026	10.00	ppbv	0.02
33) 1,4-Difluorobenzene	7.18	114	4713517	10.00	ppbv	0.02
55) Chlorobenzene-d5	12.09	117	4093340	10.00	ppbv	0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.42	95	2870438	9.70	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.00%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL092321\
Data File : VL037655.D
Acq On : 23 Sep 2021 10:22
Operator : SY/AP
Sample : VL0923ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0923ABL01

Quant Time: Sep 24 06:45:21 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL091621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Sep 16 13:42:46 2021
QLast Update : Thu Sep 16 13:42:46 2021
Response via : Initial Calibration

