

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL092321\
 Data File : VL037657.D
 Acq On : 23 Sep 2021 11:40
 Operator : SY/AP
 Sample : CAN 10024
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 CAN 10024

Manual Integrations
 APPROVED

MMDadoda
 9/27/2021 10:56:27 AM

Quant Time: Sep 24 06:48:58 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL091621AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Sep 1
 QLast Update : Thu Sep 16 13:42:46 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1529538	10.00	ppbv	0.02
33) 1,4-Difluorobenzene	7.18	114	4558757	10.00	ppbv	0.02
55) Chlorobenzene-d5	12.09	117	3951049	10.00	ppbv	0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.42	95	2749147	9.63	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.30%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
5) Vinyl Chloride	3.26	62	5314m	0.053	ppbv	
9) Propene	3.02	41	17866m	0.176	ppbv	
15) Acetone	3.89	43	25572m	0.161	ppbv	
20) Methylene Chloride	4.35	84	40641	0.536	ppbv	98
26) Hexane	5.69	57	35461	0.194	ppbv #	54
32) 1,1,1-Trichloroethane	6.52	97	13262m	0.053	ppbv	
35) Carbon Tetrachloride	7.02	117	12945m	0.050	ppbv	
38) Trichloroethene	7.83	130	7962m	0.050	ppbv	
52) Tetrachloroethene	11.22	164	8202m	0.056	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.69	83	17540m	0.052	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL092321\
Data File : VL037657.D
Acq On : 23 Sep 2021 11:40
Operator : SY/AP
Sample : CAN 10024
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CAN 10024

Manual Integrations
APPROVED

MMDadoda
9/27/2021 10:56:27 AM

Quant Time: Sep 24 06:48:58 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL091621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Sep
QLast Update : Thu Sep 16 13:42:46 2021
Response via : Initial Calibration

