

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL092321\
Data File : VL037659.D
Acq On : 23 Sep 2021 12:55
Operator : SY/AP
Sample : CAN 10334
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CAN 10334

Quant Time: Sep 24 09:49:01 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL091621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Sep 16 13:42:46 2021
QLast Update : Thu Sep 16 13:42:46 2021
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1488259	10.00	ppbv	0.02
33) 1,4-Difluorobenzene	7.18	114	4387327	10.00	ppbv	0.02
55) Chlorobenzene-d5	12.09	117	3760038	10.00	ppbv	0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.42	95	2611884	9.61	ppbv	0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.10%

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL092321\
Data File : VL037659.D
Acq On : 23 Sep 2021 12:55
Operator : SY/AP
Sample : CAN 10334
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CAN 10334

Quant Time: Sep 24 09:49:01 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL091621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Sep 16 13:42:46 2021
QLast Update : Thu Sep 16 13:42:46 2021
Response via : Initial Calibration

