

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL092418\
 Data File : VL032519.D
 Acq On : 24 Sep 2018 19:40
 Operator : SY/AP
 Sample : VSTDICC0.03
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

apatel
 9/26/2018 4:47:25 PM

Quant Time: Sep 24 21:11:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL092418AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Sep 24 2018
 QLast Update : Mon Sep 24 18:32:13 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.79	49	1525295	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.32	114	3436942	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.27	117	3426390	10.00	ppbv	-0.03

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.61	95	2588688	10.18	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.80%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.31	62	2827m	0.034	ppbv	
32) 1,1,1-Trichloroethane	6.65	97	9104m	0.044	ppbv	
35) Carbon Tetrachloride	7.16	117	9073m	0.039	ppbv	
38) Trichloroethene	8.00	130	3432m	0.031	ppbv	
52) Tetrachloroethene	11.40	164	3374m	0.037	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.87	83	11573m	0.036	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL092418\
Data File : VL032519.D
Acq On : 24 Sep 2018 19:40
Operator : SY/AP
Sample : VSTDIC0.03
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.03

Manual Integrations
APPROVED

apatel
9/26/2018 4:47:25 PM

Quant Time: Sep 24 21:11:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL092418AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Sep
QLast Update : Mon Sep 24 18:32:13 2018
Response via : Initial Calibration

