

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL092418\
 Data File : VL032529.D
 Acq On : 25 Sep 2018 4:18
 Operator : SY/AP
 Sample : J5039-02DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 OUTDOOR-AMBIENTDL

Manual Integrations
 APPROVED

apatel
 9/26/2018 4:50:34 PM

Quant Time: Sep 25 07:07:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL092418AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Tue Sep 25 07:07:56 2018
 QLast Update : Tue Sep 25 05:17:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.79	49	1410766	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.32	114	3305917	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.27	117	3308579	10.00	ppbv	-0.03

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
68) 1-Bromo-4-Fluorobenzene	14.61	95	2470090	9.90	ppbv	-0.03	
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.00%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.68	45	16930m	1.107	ppbv	
15) Acetone	3.93	43	113702	0.918	ppbv	99
20) Methylene Chloride	4.43	84	29929	0.499	ppbv	95
26) Hexane	5.80	57	35295	0.260	ppbv #	42

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL092418\
Data File : VL032529.D
Acq On : 25 Sep 2018 4:18
Operator : SY/AP
Sample : J5039-02DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
OUTDOOR-AMBIENTDL

Quant Time: Sep 25 07:07:56 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL092418AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Sep
QLast Update : Tue Sep 25 05:17:15 2018
Response via : Initial Calibration

Manual Integrations
APPROVED

apatel
9/26/2018 4:50:34 PM

